Sonicator[®] 740 Specifications

General Specifications:

ETL and C-ETL Listed:	Model ME 740 (9801427)
Input: Frequency: Input Current:	100-240 VAC ±10% 60Hz ±3 and 50Hz ±3 less than 0.6 amps
Optional Battery:	Rechargeable Smart Lithium Ion Battery Pack rated at 10.8Vand 4.8Ah
Timer Accuracy:	±10% of the indicated time
Maximum treatment time:	29 minutes
Size:	13.5 in (L) x 9.5 in (W) x 7 in (H)
Weight:	3.75 pounds without battery 4.5 pounds with battery

Ultrasonic Generator Specifications:

Frequency:	1.0 MHz ±10% 3.2 MHz ±10% 3.3 MHz ±10%
Modes:	Continuous Pulsed 10, 20 and 50% Duty cycle
Pulse repetition rate:	100 Hz ±5% (Pulse Mode)
Pulse duration:	1, 2 or 5 msec ±5%
Temporal peak/average intensity ratio:	10:1, 5:1 and 2:1 ±5%
Maximum output power:	22 W with a 10 cm² applicator, (ME 7410) 11 W with 5 cm² applicator, (ME 7413) 2.2 W with a 1 cm² applicator (ME 7431)
Maximum intensity:	2.2 W/cm ²
Indication accuracy:	$\pm 20\%$ (for any level above 10% of maximum)

Ultrasonic Applicator Specifications:

Piezoelectric discs:

The output transducer utilizes a barium titanate disc with a specially coated face.

Individual Applicator Specifications:

Applicator Part Number	Frequency	ERA
ME 7413	1 MHz ±10%	5 cm² ±20%
ME 7413	3.2 MHz ±10%	5 cm² ±20%
ME 7410	1 MHz ±10%	10 cm² ±20%
ME 7431	3.3 MHz ±10%	1 cm² ±20%

Maximum Beam Uniformity Ratio:

6:1

