
Sys*Stim[®] 208

Instruction Manual



1333 South Claudina Street • Anaheim, CA 92805, U. S. A.
Toll Free: (800) 854-9305 • Telephone: (714) 533-2221 • FAX: (714) 635-7539
Web Site: www.mettlerelectronics.com • Email: mail@mettlerelectronics.com

IR9-23

June, 2006

Copyright © 2003 by Mettler Electronics Corp.—Anaheim, CA

Table of Contents

| Section | Title | Page |
|----------------|--|-------------|
| 1 | Introduction | 5 |
| 1.1 | Introduction to the Sys*Stim 208 | 5 |
| 1.2 | Introduction to this Manual | 5 |
| 1.3 | Safety Precautions | 6 |
| 1.4 | Caution | 6 |
| 1.5 | Shipping Damage | 6 |
| 1.6 | Package Contents | 7 |
| 1.7 | Limited Warranty | 7 |
| 2 | Control Descriptions and List of Abbreviations | 9 |
| 2.1 | Control Descriptions | 9 |
| 2.2 | List of Abbreviations | 10 |
| 3 | Installation Instructions | 11 |
| 4 | Operating Instructions | 13 |
| 4.1 | A Note About Electrodes | 13 |
| 4.2 | General Operating Instructions | 14 |
| 4.3 | Operating Instructions | 14 |
| 5 | Indications, Contraindications, Precautions and Adverse Reactions | 17 |
| 5.1 | Indications | 17 |
| 5.2 | Contraindications | 17 |
| 5.3 | Warnings | 18 |
| 5.4 | Precautions | 19 |
| 5.5 | Side Effects/Adverse Reactions | 19 |
| 6 | Maintenance and Troubleshooting | 21 |
| 6.1 | Cleaning the Sys*Stim 208 | 21 |
| 6.2 | Routine Maintenance | 21 |
| 6.3 | Changing Fuses | 21 |
| 6.4 | Troubleshooting the Sys*Stim 208 | 22 |
| 7 | Specifications | 23 |
| 7.1 | General Specifications | 23 |
| 7.2 | Output Specifications | 23 |
| 8 | Accessories | 25 |
| 8.1 | Ordering Information | 25 |
| 8.2 | Sys*Stim 208 Accessories | 25 |

List of Figures

| No. | Title | Page |
|------------|---|-------------|
| 1.1 | Sys*Stim 208 | 5 |
| 3.1 | Sys*Stim 208, Back view—Mains Power Switch and Line Cord connection | 11 |
| 3.2 | Sys*Stim 208, Front View—Electrode Cable Connections | 11 |
| 4.1 | Front membrane panel and LED indicators | 13 |
| 4.2 | Electrode Sizes and Current Density | 14 |
| 7.1 | Waveform Illustration | 23 |

Section 1: Introduction

1.1 Introduction to the Sys*Stim 208

Thank you for purchasing the Sys*Stim 208 one-channel Low Volt neuromuscular stimulator. The microprocessor controlled Sys*Stim 208 produces low volt current through one channel. The unit produces an asymmetrical electrically balanced waveform. There are three modes of operation: Pulse—1 to 80 Hz, Tetanize—80 Hz and Surge—80 Hz, On/Off times variable.

The Surge mode produces an On/Off time from 0.5 to 3.75 seconds. This mode is used to contract and relax muscles.

The Sys*Stim 208 is portable and beautifully designed. Up and down buttons control the timer while easy-to-use knobs allow you to select treatment parameters and adjust intensity.

An accessory for the Sys*Stim 208 is the Patient Safety Switch, which is connected to the jack located on back of the unit and handed to the patient during treatment so that they can stop treatment if it becomes uncomfortable.



Figure 1.1— Sys*Stim 208

1.2 Introduction to This Manual

Read the contents of this manual before treating patients with the Sys*Stim 208.

This manual has been written to assist you with the safe operation of the Sys*Stim 208. It is intended for use by the owners and operators of the

Sys*Stim 208. The goal of this manual is to direct the correct operation and maintenance of this unit.

The specifications and instructions presented in this manual are in effect at the time of its publication. These instructions may be updated at any time at the discretion of the manufacturer.

1.3 Safety Precautions

The Sys*Stim 208 operates with high voltages. Only qualified biomedical technicians with training in neuromuscular stimulator service should perform servicing of the Sys*Stim 208 or it should be returned directly to the factory. To maximize safety during use, the unit should be plugged into a grounded wall outlet. General safety guidelines for medical electronic equipment should be followed.

Service may be obtained from the manufacturer by sending the Sys*Stim 208 in its original shipping container to Mettler Electronics Corp., 1333 South Claudina Street, Anaheim, CA 92805, ATTN: Service Department. (Telephone toll free: (800) 854-9305, Email: service@mettlerelectronics.com, Alternate telephone number: 1 (714) 533-2221)

NOTE: All warranty repairs must be performed by Mettler Electronics Corp. or by a service facility authorized by Mettler Electronics to perform warranty repair work.

A service manual for the Sys*Stim 208 is available from Mettler Electronics Corp. for a nominal charge.

1.4 Caution

Federal law restricts the sale of this device to, or on the order of, a physician, dentist, veterinarian or any other practitioner licensed by law of the state in which he practices.

Use of controls or adjustments or performance of procedures other than those specified herein may result in hazardous exposure to electrical energy. The electric energy delivered by this device may possibly be lethal. Treatment should be administered only under the direct supervision of a health care professional.

1.5 Shipping Damage

Your new Sys*Stim 208 is shipped complete in one carton. Upon receipt, please inspect the carton and the unit for visible and hidden damage. If you discover any damage, hold all shipping materials, including the carton, and call the shipping agent who delivered the unit. They are responsible for all damage in transit; therefore, all claims should be filed directly with them. The factory will not be responsible for any damage in shipment, nor

allow any adjustments unless proper formal claim has been filed by the receiver against the carrier.

The carton in which your new Sys*Stim 208 was received is specially designed to protect the unit during shipping. Please retain all shipping materials in the event that you will need to return your unit for servicing.

1.6 Package Contents

Your new Sys*Stim 208 comes complete with all the necessary components to perform neuromuscular electrical stimulation. Below is a list of items that are included in the shipping carton.

1. Sys*Stim 208
2. One electrode cable set, (ME 2260)
3. Two gray pin to banana adapters, (ME 2027)
4. One package V Trodes, 2" diameter (ME 2702)
5. One Patient Safety Switch, (ME 2031)
6. Detachable U.L. listed, hospital-grade line cord, (ME 7293)
7. Instruction Manual and Warranty Card

1.7 Limited Warranty

The Sys*Stim 208 neuromuscular electrical stimulator is warranted against defects in materials and workmanship for a period of two years from date of purchase. During the applicable warranty period Mettler Electronics Corp. will, at its discretion, either repair or replace the Product without charge for these types of defects.

For service under this warranty, the Product must be returned by the buyer within the applicable warranty period to Mettler Electronics Corp.

Shipping charges to Mettler Electronics Corp. under this warranty must be paid by the buyer. The buyer must also include a copy of the sales receipt or other proof of the date of purchase. If the Product is returned without proof of the date of purchase, it will be serviced as an out-of-warranty product at Mettler Electronics Corp.'s prevailing service rates.

Alteration, misuse, or neglect of the Product voids this warranty. Except as specifically set forth above, Mettler Electronics Corp. makes no warranties, express or implied, including without limitation any implied warranty of merchantability or fitness for a particular purpose, with respect to the Product. If any implied warranties apply as a matter of law, they are limited in duration to one year.

Mettler Electronics Corp. shall not be liable for any indirect, special, consequential or incidental damages resulting from any defect in or use of the Product.

Any legal action brought by the buyer relating to this warranty must be commenced within one year from the date any claim arises and must be brought only in the state or federal courts located in Orange County, California.


Some states do not allow limitations on how long an implied warranty lasts, or the exclusion or limitation of incidental or consequential damages, so the above limitations or exclusions may not apply to the buyer. This warranty gives the buyer specific legal rights, and the buyer may also have other rights which vary from state to state.

Section 2—Control Descriptions and List of Abbreviations

2.1 Control Descriptions




Time display shows time remaining in treatment in whole minutes. When the unit is first turned on the time display will show 15 minutes. You can set any time from 1-60 minutes.




The timer controls allow you to adjust treatment time up or down.



The Reset indicator turns off when the intensity control is fully rotated to the "Off" position.




The Rate control knob adjusts pulse frequency and Surge On/Off times.



The Mode control knob allows you to select the treatment mode.



The electrode cable is plugged into the electrode jack.



The Intensity Control allows you to adjust treatment intensity. The indicator light will illuminate when stimulation output is active.



Mains Off.



Mains On.



Type BF Equipment—Class I



Attention, consult instruction manual.

2.2 List of Abbreviations

| | | |
|---------------|---|--|
| Hz | — | Hertz (pulses per second) |
| LED | — | Light Emitting Diode |
| μC | — | microcoulombs |
| μs | — | Microsecond (1×10^{-6} second) |
| mA | — | Milliampere (1×10^{-3} ampere) |
| min | — | Minutes |
| s | — | Seconds |
| S/N | — | Serial Number |
| V | — | Volts |

Section 3—Installation

3.1 Installation Instructions

1. Connect the line cord to the back of the Sys*Stim 208. (See Figure 3.1)
2. Plug the line cord (ME 7293) into a grounded wall outlet that is rated at 100 to 240 Volt AC 50/60 Hz. Your power supply must match the voltage requirements listed on the serial number label of your device.
Do not connect the Sys*Stim 208 to a power supply rated differently than that described above.
3. The line cord comes equipped with a standard 3-prong plug. This plug provides grounding for the Sys*Stim 208. Do not defeat its purpose by using 3-to-2 prong adapters or any other means of attaching to a wall outlet.
4. Plug the electrode cable (ME 2260) into the electrode cable jack as seen in Figure 3.2.
5. Plug the Patient Safety Switch (ME2031) into the back of the unit as seen in Figure 3.1.
6. The Sys*Stim 208 may be susceptible to interference originating from shortwave diathermy units operating in close proximity to it. Avoid operating the Sys*Stim 208 adjacent to and simultaneously with operating shortwave devices.
7. Once you have verified proper functioning of your Sys*Stim 208, using the instructions in Section 4, please fill in the enclosed self-addressed Warranty Registration Card and mail it to Mettler Electronics.




Figure 3.1— Sys*Stim 208, Back View—Mains Power Switch, Line Cord Connection and Patient Safety Switch Connection




Figure 3.2— Sys*Stim 208, Front View—Electrode Cable Connection

Section 4—Operating Instructions




Figure 4.1—Front membrane panel and LED indicators

4.1 A Note about Electrodes

To ensure safe operation of the Sys*Stim 208, follow the recommendations listed below:

1. We strongly encourage careful maintenance of the electrode system. This includes the lead wires as well as the pads themselves. Worn cables and/or poor pads (or the wrong sized pads) can have a significant impact upon treatment results.
2. Do not exceed the number of recommended uses listed on the instructions for V Trodes or other reusable self-adhesive electrodes.
3. Make sure that the entire surface of the electrode is contacting the patient.
4. Do not use moist hot packs to secure electrodes.
5. To avoid skin irritation due to high current density, do not use electrodes smaller in surface area than the 2" diameter V Trode™ self-adhesive electrode (ME 2702).
6. Do not use conductive carbon electrodes with this product.
7. Whenever clinically possible, utilize the largest possible pads to reduce local increases in current density. In situations where small pads are required, use the lowest stimulation intensity necessary to achieve the desired clinical results.

The table below illustrates the relationship between electrode diameter and current density. As you can see, the current density increases rapidly when diameter decreases.

| Diameter <i>inches</i> | Surface Area <i>Square inches</i> | Current Density <i>mA/sq in (for 10mA)</i> |
|----------------------------------|---|--|
| 1.25 | 1.2 | 8.2 |
| 2.00 | 3.1 | 3.2 |
| 3.00 | 7.1 | 1.4 |





Figure 4.2—Electrode Sizes and Current Density

4.2 General Operating Instructions:


Before you start.

- Review precautions, contraindications and side effects/adverse reactions listed in Section 5.
- Use Mettler Electronics electrodes to ensure safe and effective operation.
- Verify connection of the line cord to a grounded wall receptacle and the Sys*Stim 208.
- For electrical stimulation connect the electrode cable (ME 2260) into the electrode connection.
- Connect the Patient Safety Switch (ME2031) into the back of the unit and hand it to the patient.
- Note: Descriptions of the symbols used on controls are in Section 2.


4.3 Operating Instructions




- Turn on the mains power switch by pressing up on the On/Off switch located on the back of the unit.



2. Apply the electrodes to the patient. Attach the electrode cable to the electrodes.




3. Reset the intensity control by rotating it fully counterclockwise to the "Off" position. The Reset LED indicator should turn off.




4. Select Treatment Mode:

- Pulse, 1-80 Hz
- Tetanize, 80 Hz
- Surge, 80 Hz, 0.5 s to 3.75 s On/Off



5. Set the Rate by turning the control knob clockwise to increase and counterclockwise to decrease:

- Pulse, 1-80 Hz
- Surge, 0.5 s to 3.75 s On/Off



6. Set the Timer by holding down the up arrow to increase the time and the down arrow to decrease the time. The default time is 15 minutes.

7. Adjust the stimulation intensity to patient tolerance by turning the intensity control knob clockwise to increase intensity or counterclockwise to decrease intensity.



8. At the end of the treatment the timer returns to zero and the unit will beep.

Section 5—Indications, Contraindications, Precautions and Adverse Reactions

5.1 Indications

The application of pulsating electric currents to the body via electrodes elicits responses from nerves, which conduct pain sensation and muscle contraction information. Stimulation of sensory fibers will help block pain while the stimulation of motor fibers will generate pulsatile contractions of the muscle groups innervated by the nerves being stimulated.

Based on this information, some of the indications for use are as follow:

1. Symptomatic relief of chronic intractable pain, acute post traumatic pain or acute post surgical pain.
2. Temporary relaxation of muscle spasm.
3. Prevention of post–surgical phlebo–thrombosis through immediate stimulation of calf muscles.
4. Increasing local blood circulation.
5. Prevention or retardation of disuse atrophy.
6. Muscle re–education.
7. Maintaining or increasing range of motion.

5.2 Contraindications

1. Electrical neuromuscular stimulation should not be administered to individuals who are or may be pregnant.
2. Do not stimulate a patient who has a cardiac demand pacemaker.
3. Patients with implanted electronic devices should not be subjected to stimulation.
4. Placement of electrodes across the chest laterally or anterior/posterior creates a possible hazard with cardiac patients and is therefore not recommended. Do not use transthoracically in any mode. Great care should be exercised in applying the electrical stimulus current to any region of the thorax because the stimulus current may produce cardiac arrhythmia. In patients with known heart disease, electrical stimulation should be used only after careful physician evaluation and patient instruction.
5. Place electrodes in such a way to avoid stimulation of the carotid sinus (neck) region.

6. Patients with arterial or venous thrombosis, or thrombophlebitis are at risk of developing embolisms when electrical stimulation is applied over or adjacent to the vessels containing the thrombus. If a patient has a history of deep vein thrombosis, even many years past, the affected area should not be stimulated.
7. Do not use over swollen, infected, or inflamed areas. Do not place electrodes over skin eruptions.
8. Fresh fractures should not be stimulated in order to avoid unwanted motion.
9. Do not apply stimulation transcerebrally (through the head).
10. Do not use on cancer patients.
11. Stimulation should not be applied immediately following trauma or to tissues susceptible to hemorrhage.
12. Positioning electrodes over the neck or mouth may cause severe spasm of the laryngeal or pharyngeal muscles. These contractions may be strong enough to close the airway or cause difficulty in breathing.
13. Do not apply stimulation for undiagnosed pain syndromes, until etiology is established.
14. Do not apply electrodes directly over the eyes or inside body cavities.
15. Do not use electrical stimulation in conjunction with high frequency surgical equipment or microwave therapy systems.

5.3 Warnings

1. Electrical stimulation is ineffective for pain of central origin, this includes headache.
2. Electrical stimulation must be applied by a physician or other qualified practitioner and should be used for only the prescribed purposes.
3. Electrical stimulation is of no curative value.
4. Electrical stimulation is a symptomatic treatment and as such suppresses the sensation of pain, which could serve as a protective mechanism.
5. The safety of electrical stimulators for use on children has not been determined. Keep out of reach of children.
6. The long-term effects of chronic electrical stimulation are unknown.
7. Electronic monitoring equipment (such as ECG monitors and ECG alarms) may not operate properly when electrical stimulation is in use.

5.4 Precautions

1. Care should be taken in the treatment of patients receiving another type of electrotherapeutic treatment (such as conventional TENS) or having indwelling electrodes, lead wires, or transmitters (for electrophrenic pacing or cerebellar or urinary bladder stimulation). Stimulation currents should not cross the lead wires or electrodes.
2. It is advisable to insulate patients, preferably by use of a wooden treatment table or one that is completely padded by non-conductive material. Added safety is provided if the patient cannot touch any grounded metal parts.
3. To prevent burns, avoid current densities exceeding 2 mA/cm² when using this device.
4. Isolated cases of skin irritation may occur at the site of electrode placement following long-term application.
5. Avoid placing electrodes directly over open wounds since current density tends to concentrate in these areas.
6. Use extreme caution when treating desensitized areas or on patients who may not be able to report discomfort or pain.
7. Use caution in applying electrical stimulation over areas where there is a loss of normal skin sensation.
8. Adequate precautions should be taken in the case of persons with suspected or diagnosed epilepsy.
9. Never leave the patient unattended during treatment without placing the patient treatment safety switch within the patient's reach and instructing the patient how to use it.
10. Care should be taken following recent surgical procedures when muscle contraction may disrupt the healing process.
11. Do not apply electrical stimulation over the menstruating uterus.
12. Effectiveness for pain management is highly dependent upon patient selection by a person qualified in the management of pain patients.
13. Electrode placement and stimulation settings should only be based on the guidance of the prescribing practitioner.
14. Electrical stimulators should be used only with the leads and electrodes recommended for use by the manufacturer.
15. Turn on the Sys*Stim 208 before applying electrodes to the patient.

5.5 Side Effects/Adverse Reactions

1. Skin irritation and burns beneath the electrodes have been reported with the use of electrical muscle stimulators.
2. Possible allergic reactions to tape, gel or electrodes may occur.

Section 6—Maintenance and Troubleshooting

6.1 Cleaning the Sys*Stim 208

1. The Sys*Stim 208 can be wiped off with a damp cloth. The power cord should be disconnected from the unit before this is done. In the case of stubborn dirt a gentle household cleaner can be sprayed on the cloth and then wiped on the unit. If this method is used, remove any cleaner residue with a damp cloth.
2. Follow the V Trode package insert for the use and care of the electrodes supplied with the Sys*Stim 208.
3. For routine cleaning of the electrode cables use soap and water. Thoroughly dry after cleaning.

6.2 Routine Maintenance

1. Standard medical electrical safety checks should be performed annually by qualified biomedical engineers or technicians trained to perform these procedures.
2. Inspect electrode cable and associated connector for damage.

6.3 Changing Fuses

Follow the sequence below when changing fuses:

1. Unplug the power cord from the wall and the back of the unit.
2. Pull out fuse drawer. The Fuse Block is located at below the Mains Plug on the unit. Pull it out until it comes completely out of the unit.
3. Remove fuse located at the rear of the fuse block. Please note: The fuse in the hollow tube of the fuse block is a spare.
4. Replace with a fuse of the same type and value.
5. Place fuse block back into Power Inlet.

6.4 Troubleshooting the Sys*Stim 208

Symptom

1. Nothing lights when main power switch is turned on.

2. There is no stimulation output.

Action

Is line cord connected to outlet?
Does the outlet have power?
Unit may require a new fuse. See Section 6.3 for replacing the fuse.

Turn intensity control to the “Off” position. The “Reset” indicator should go out. If there is still no output, check the lead wires and electrodes for damage. If none of these actions is successful, the unit may require servicing.

If problem is not addressed above, or if additional troubleshooting guidance is desired, call (800) 854-9305 or email service directly at service@mettlerelectronics.com.

Section 7—Specifications

7.1 General Specifications:

| | |
|-------------------------|--|
| Input: | 100 to 240 Volt AC 50/60 Hz, 0.75 Amp (max) |
| External Fuse: | 0.75 A, 250 V, Slow Blow |
| Weight: | 1.4 pounds |
| Dimensions: | 2.75 in (H) x 6.1 in (D) x 8 in (L) |
| Maximum Treatment Time: | 60 minutes |

7.2 Output Specifications:

| | |
|-----------------|---|
| Channels: | One |
| Waveform: | Asymmetrical biphasic with zero net DC |
| Voltage: | 110 V peak into a 1k ohm load 28 V peak into a 100 ohm load |
| Intensity: | 56 μ C Max per Pulse into a 100 ohm load |
| Phase Duration: | 200 μ s @ 50% Max Amplitude |
| Frequency: | Pulse Mode: 1-80 Hz Tetanzize Mode: 80 Hz Surge Mode: 80 Hz |
| Surge Cycle: | 0.5 to 3.75 seconds On/Off |




Figure 7.1—Waveform Illustration

Section 8—Accessories

8.1 Ordering Information:

Therapy products and accessories are available from Mettler Electronics authorized Distributors. For information regarding either Mettler products or a distributor near you, please call toll free, (800) 854–9305 or phone (714) 533–2221 in areas outside the continental United States. Ask for Customer Service. Mettler Electronics is open from 7 AM until 5 PM Pacific Time for your convenience. The email address for Customer Service is mail@mettlerelectronics.com.

8.2 Sys*Stim 208 Accessories

| Catalogue # | Item Description |
|-------------|---|
| 2000 | 4 Sponge electrodes (2" x 2") |
| 2001 | 24 Sponge inserts (2" x 2") |
| 2002 | 4 Sponge electrodes (4" x 4") |
| 2003 | 24 Sponge inserts (4" x 4") |
| 2004 | 1 Sponge electrode (3.5" x 7") |
| 2005 | 12 Sponge inserts (3.5" x 7") |
| 2006 | 1 Sponge electrode (8" x 10") |
| 2007 | 12 Sponge inserts (8" x 10") |
| 2008 | 4 Electrode straps (24") |
| 2009 | 4 Electrode straps (48") |
| 2023 | Pencil electrode set with push button stimulation control, (includes handle, 4 different sizes of stainless steel spot electrode tips, and carrying case) |
| 2027 | Pin to banana adapter plug set to be used with ME 2026, 2260 or 2201 electrode cables. Four each, gray. |
| 2030 | Bifurcated cord set, one red and one black, , pin termination |
| 2031 | Patient Safety Switch |
| 2221 | EZ Trode – 2" diameter round self–adhering, reusable electrodes with lead wires; case of ten packages (four electrodes/pkg.) |
| 2222 | EZ Trode – 2.75" diameter round self–adhering, reusable electrodes with lead wires; case of ten packages (four electrodes/pkg.) |
| 2223 | EZ Trode – 2" x 5" self–adhering, reusable electrodes with lead wires, case of 10 packages (2 electrodes/pkg.) |
| 2224 | EZ Trode – 2" square self–adhering, reusable |

- electrodes with lead wires; case of ten packages (four electrodes/pkg.)
- 2260 Electrode cable for the Sys*Stim 208 with pins
- 2702 V Trode –2" diameter round electrodes with lead wires, case of ten packages (four electrodes/pkg.)
- 2703 V Trode –2.75" diameter round electrodes with lead wires, case of 10 packages (four electrodes/pkg.)
- 2704 V Trode –2" x 4" oval electrodes with lead wires, case of 10 packages (four electrodes/pkg.)
- 2705 V Trode –2" square electrodes with lead wires, case of 10 packages (four electrodes/pkg.)
- 7293 Detachable U.L. listed, hospital–grade line cord